			De et II.	
MALLORY Mallory Sonalert Products,Inc.			Part #: SBT12MMPPC	
Sales Outline Drawing		Revision:		
Specifications:	cUL Recognized to IEC60601-1-8 3rd Edition Amendment 1			
	Artificial Perfusion Melody			
Sound Level Category:	Silence	Loud	Loud	Loud
Mode of Operation:	Silence - NO SOUND - See Note 1	Low Priority Alarm - 2 beeps repeating every 20 sec	Medium Priority Alarm: 3 beeps repeating every 7.5 sec	High Priority Alarm - 10 beeps repeating every 2.5 sec
Loudness @ 10cm:	Silence	85 to 95 dB(A) Typ.	85 to 95 dB(A) Typ.	
Frequency:	Silence	150 to 4000 Hz	150 to 4000 Hz	150 to 4000 Hz
Connection:	Connect "+" to "+" & "-" to "-" of Power Supply	Connect "+" to "+" & "-"& L to "-" of Power Supply	Connect "+" to "+" & "-" & M to "-" of Power Supply	Connect "+" to "+" & "-" & H to "-" of Power Supply
Voltage Rating:		9 to 12 \		
Average Current Draw:	<15mA Typ.	<15mA Typ.	<15mA Typ.	<20mA Typ.
Ipk Current*:	75mA			
Housing Material:	ABS 777D Plastic Resin, Black			
Storage Temperature:	-40° to +85° C			
Operating Temperature:	-30° to +65° C			
Weight Typical:	0.6 oz (17g)			
Options: Condition by wave solder (°C)	Please contact factory. 260 ± 5 within 5 sec			
Condition by wave solder (°C)	350 ± 20 within 5 sec			
	cUL Recognized ROHS Compliant			
Dimensions: Inches [mm]	PATENT NUMBER: 7880593			
PATEINT INUIVIDER. 70000995				
Ø1.750 [Ø44.5] 0.334 [8.5]				
	0.100 [2.5 0.551 [14.0]	0.200 [5.1]		NAME OF THE PARTY

Note 1: Not inteded to be energized when alarm not in use.

*Ipk = Instantaneous Peak

tolerances:

±0.01 [±0.3]

±0.005 [±0.1]